

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

NOV 20 2009

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130114
Name Kevin Stevenson ID# 2414 BMB Date 12/31/09
12414 Date 11-14-09

A Agency WAA Police Dept Phone # 786-1120

Instrument Location 2601 Providence Dr. Suite 114 ESH

B Alco S/N 75470 Target Value .080 High Pressure 800

Alco Test Value Average .079 .078
1st Alco 2nd Alco

Signature [Signature]

(OVER)

BMB
12/31/09


(Do Not write in the area below)

- I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
 - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
 - (3) I am the Scientific Director of the State Breath Alcohol Program.
 - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
 - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
 - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 31st day of Dec., 2009.

[Signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130114

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130114

NOVEMBER 14, 2009

OPERATOR'S NAME:
STEVENSON KEVIN L
OPERATOR'S NUMBER: 2414

SUBJECT'S LAST NAME:
VERIFICATION
SUBJECT'S FIRST NAME/MI :
VERIFICATION

D.L. #: 1234
DEPT/AGENCY: ANC7
CASE/REPORT: 090876

TEST TYPE: U
ALCO TARGET VALUE: .000
ALCO S/N: 75470

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 28.86 in
ALCO TARGET .077 17:42
BLANK TEST .000 17:43
INTERNAL STANDARD VERIFIED 17:43
ALCO TO 28.86 in .079 17:44
BLANK TEST .000 17:44
SUBJECT SAMPLE .000 17:45
BLANK TEST .000 17:46
ALCO TO 28.86 in .078 17:46
BLANK TEST .000 17:47

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130114

NOVEMBER 14, 2009
TIME 17:39

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 28.84 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!""%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxy z{|}~